EU DECLARATION OF CONFORMITY

MANUFACTURER:

Innovative Scientific Solutions, Inc.

7610 McEwen Road Dayton, OH 45459 U.S.A.

Phone: 937-630-3012, Fax: 937-630-3015

URL: www.innssi.com

AUTHORISED

Neil Stokes

REPRESENTATIVE:

Optical Measurement Systems Manager

Aircraft Research Association Ltd

Manton Lane Bedford England MK41 7PF

Direct Dial Tel: +44(0) 1234 324661 Switchboard: +44(0) 1234 324600

Email: nstokes@ara.co.uk
Website: www.ara.co.uk

MODEL / TYPE:

LC-1s CCTV Lens Controller

DESCRIPTION:

Remote controller for CCTV Motorized and Auto Iris Lenses

SERIAL NUMBER:

TBD

REPORTS:

STF.2019.ISSI.1

DIRECTIVES:

Low Voltage Directive 2014/35/EU, EMC Directive 2014/30/EU, RoHS Directive 2011/65/EU

STANDARDS CONSIDERED, FULL OR IN PART:

- -EN 61010-1:2010 -Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements
- EN 55011:2009+A1:2010 Industrial, scientific, and medical equipment Radio-frequency disturbance characteristics Limits and methods of measurement
- -EN 61326-1:2013 Electrical equipment for measurement, control, and laboratory use EMC requirements Part 1: General requirements
- -EN 61000-3-2:2006, Inc. A1:2009 & A2:2009 Electromagnetic compatibility (EMC) -Part 3-2: Limits Limits for harmonic current emissions (equipment input current ≤ 16A per phase)
- -EN 61000-3-3:2013 Electromagnetic compatibility (EMC) -Part 3-3: Limits Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16A per phase and not subject to conditional connection
- -EN 50581:2012 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

conformity with the Low Voltage Directive 2014/35/EU, the EMO	Directive 2014/30/EU, and the RoHS Directive 2011/65/EU
Authorized by: 2 Lln	Date: 7/7/20
Name: Elic Graham	Title: Production Manager
Location: ISSI HQ	β*

Innovative Scientific Solutions, Inc. declares under its sole responsibility that the CCTV Lens Controller, model LC-1s is in